

Certificate for Experimental Results

It is certified that the result of test on inactivation of bacteria as below;

1. Manufacturer: INB Co.,Ltd (President: Lee, Kang-Soo)
#522 Unitechvil, Baksuk-dong, Ilsan-gu, Koyang-si, Kyunggi-do, Korea
2. Product name: VirusKiller
3. Strain and its titer used for inactivation:
 - Klebsiella pneumoniae KCTC 2241
 - Staphylococcus aureus subsp. aureus KCTC 1928
 - Streptococcus pyogenes KCTC 3096
 - Streptococcus pneumoniae CP 1200
 - Escherichia coli DH 5
4. Experimental methods: Number of surviving bacteria obtained from filters captured exhaust air, after direct introduction of bacteria into the air cleaner(VirusKiller) by aerosol spray with 20mL solution containing the above titer of bacteria.
5. Experimental result: No surviving bacteria from the exhaust air was detected.
The air cleaner(VirusKiller) inactivated all of the bacteria added for the test.

April 19, 2004



Prof. Zee, Ok-Pyo, Ph. D.

**Director of Kyunggi Pharmaceutical Research Center
of Sungkyunkwan University**